

## **Certificate of Analysis**

Product	SULFURIC ACID, 0.1N	Batch No.	: 21 05 0227
Code No.	GN1194	Manufacturing date	: 19/05/2021
		Best before	: 05/2023

Items	Test	Specifications	Test Results
1	Appearance	Clear, colorless solution	Passes test
2	Normality	0.1000N $\pm$ 0.0005N	0.1001N

Traceable to NIST : 723e Tris Acidimetric

This is to certify the above batch has been tested and complies with this specification.

Ladaphan.

Chemist

Achan .

Technical Manager