RCI Labscan

RCI Labscan Limited

Certificate of Analysis

Product	**HYDROCHLORIC ACID, 0.1N**	**Batch No.**	**:**	**21 06 0178**
Code No.	**GN1108**	**Manufacturing date**	**:**	**24/06/2021**
		Best before	**:**	**06/2023**

Items	Test	Specifications	Test Results
1	Appearance	Clear, colorless solution	Passes test
2	Normality	0.1000N ± 0.0005N	0.1001N

Traceable to NIST : 723e Tris Acidimetric

This is to certify the above batch has been tested and complies with this specification.

Ladaphan

Chemist

Technical Manager