

Certificate of Analysis

Product	BUFFER SOLUTION pH 7.0 (Green)	Batch No.	: 21 06 0322
Code No.	GN1302	Test date	: 30/06/2021
		Best before	: 06/2023

Items	Test	Specifications	Test Results
1	Appearance	Clear, green solution	Passes test
2	pH at 25°C	pH 7.00 ± 0.02	pH 7.00

Traceable to NIST SRMs : 186g

This is to certify the above batch has been tested and complies with this specification.

Deviations of pH (Δ pH) at various temperatures :

	Δ pH
5 °C	+ 0.04
10 °C	+ 0.02
15 °C	+ 0.01
20 °C	+ 0.01
25 °C	± 0
30 °C	- 0.01
35 °C	- 0.01
40 °C	- 0.02
50 °C	- 0.02

Ladaphan.

Chemist

Achan.

Technical Manager