

Certificate of Analysis

Product	HYDROCHLORIC ACID, 1.0N	Batch No.	: 25 10 0366
Code No.	GN1111	Manufacturing date	: 14/10/2025
		Best before	: 10/2027

Items	Test	Specifications	Test Results
1	Appearance	Clear, colorless solution	Passes test
2	Normality	1.000N $\pm$ 0.005N	1.002N

Traceable to NIST : 723e Tris Acidimetric

This is to certify the above batch has been tested and complies with this specification.

Ladaphan
Chemist

Achan .
Technical Manager